

ReSound ONE

User guide

ReSound Receiver-In-Ear hearing aids

Hearing aid information

| Left hearing aid | | | Right hearing aid | | | | | |
|---------------------|-----------|--|------------------------------|-------------------------------|------------|-----|---------------|--|
| Serial number | | | | Serial number | | | | |
| Model number | | | Model number | | | | | |
| Battery type ☐ Rech | | nargeable 🗆 Zinc-Air 312 🗆 Zinc-Air 13 | | | | | | |
| Domo/ cormold | Classed d | omo | Onon domo | Dawar dama | □ Tulin de | | □ DIE cormold | |
| Dome/ earmold type | □ Small | | Open dome ☐ Small ☐ Medium | Power dome ☐ Small ☐ Medium | ☐ Tulip do | ome | □ RIE earmold | |
| | □Large | | □Large | □ Large | | | | |
| Program | | Веер | | Description | | | | |
| 1 | | One beep | | | | | | |
| 2 | | Two beeps | | | | | | |
| 3 | | Three beeps | | | | | | |
| Δ | | Four beens | | | | | | |

FDA warnings and cautions (US only)

WARNING: People younger than 18 should go to a doctor before using this

People younger than 18 years old need specialized care, and using this without a medical evaluation may worsen impairment or disability. A hearing aid user who is younger than 18 should have a recent medical evaluation from a doctor, preferably an ear-nose-throat doctor (an ENT). Before using this, a doctor should determine that the use of a hearing aid is appropriate.



WARNINGS to Hearing Aid Dispensers

You should advise a prospective hearing aid user to consult promptly with a doctor, preferably an ear specialist such as an ENT, before dispensing a hearing aid if you determine through inquiry, actual observation, or review of any other available information concerning the prospective user, that the prospective user has any of the following conditions:

- Visible deformity of the ear, either congenital or traumatic
- Fluid, pus, or blood coming out of the ear within the previous 6 months
- Pain or discomfort in the ear
- History of excessive ear wax or suspicion that something is in the ear canal
- Dizziness, either recent or long-standing
- Sudden, quickly worsening, or fluctuating hearing loss within the previous 6 months

- · Hearing loss or ringing (tinnitus) only in one ear or noticeable difference in hearing between ears
- Audiometric air-bone gap equal to or greater than 15 dB at 500 Hz, 1000 Hz, and 2000 Hz.

Outputs over 132 dB SPL:

You should exercise special care in selecting and fitting a hearing aid with a maximum output that exceeds 132 dB SPL because it may impair the remaining hearing of the hearing aid user.

Sound pressure level in the ears of children:

The developed sound pressure level in the ears of children can be substantially higher than in average adults. It is recommended to perform an RECD measurement in order to ensure the correct target for the fitted OSPL90.



CAUTION:

This is not hearing protection

You should remove this device if you experience overly loud sounds, whether short or long-lasting. If you're in a loud place, you should use the right kind of hearing protection instead of wearing this device. In general, if you would use ear plugs in a loud place, you should remove this device and use ear plugs.

The sound output should not be uncomfortable or painful

You should turn down the volume or remove the device if the sound output is uncomfortably loud or painful. If you consistently need to turn the volume down, you may need to further adjust your device.

You might need medical help if a piece gets stuck in your ear

If any part of your hearing aid, like the eartip (dome), gets stuck in your ear, and you can't easily remove it with your fingers, get medical help as soon as you can. You should not try to use tweezers or cotton swabs because they can push the part further into your ear, injuring your eardrum or ear canal, possibly seriously.



What you might expect when you start using your hearing aid

- · A hearing aid can benefit many people with hearing loss. However, you should know it will not restore normal hearing, and you may still have some difficulty hearing over noise. Further, a hearing aid will not prevent or improve a medical condition that causes hearing loss.
- People who start using hearing aids sometimes need a few weeks to get used to them. Similarly, many people find that training or counseling can help them get more out of their devices.
- If you have hearing loss in both ears, you might get more out of using hearing aids in both, especially in situations that make you tired from listening – for example, noisy environments.

Tell FDA about injuries, malfunctions, or other adverse events

To report a problem involving your hearing aid, you should submit information to FDA as soon as possible after the problem. FDA calls them "adverse events", and they might include: skin irritation in your ear, injury from the device (like cuts or scratches, or burns from an overheated battery), pieces of the device getting stuck in your ear, suddenly worsening hearing loss from using the device, etc. Instructions for reporting are available at https://www.fda.gov/Safety/MedWatch. or call 1-800-FDA-1088 You can also download a form to mail to FDA

Hearing loss in people younger than 18

- People vounger than 18 should see a doctor first, preferably an ear-nose-throat doctor (an ENT). because they may have different needs than adults.
- The doctor will identify and treat medical conditions as appropriate.
- The doctor may refer the person to an audiologist for a separate test, a hearing aid evaluation.
- The hearing aid evaluation will help the audiologist select and fit the appropriate hearing aid.

A person who is younger than 18 years old with hearing loss should have a medical evaluation by a doctor, preferably an ENT, before buying a hearing aid. The purpose of a medical evaluation is to identify and treat medical conditions that may affect hearing but that a hearing aid won't treat on its own.

Following the medical evaluation and if appropriate, the doctor will provide a written statement that the hearing loss has been medically evaluated and the person is a candidate for a hearing aid. The doctor may refer the person to an audiologist for a hearing aid evaluation, which is different from the medical evaluation and is intended to identify the appropriate hearing aid.

The audiologist will conduct a hearing aid evaluation to assess the person's ability to hear with and without a hearing aid. This will enable the audiologist to select and fit a hearing aid for the person's individual needs. An audiologist can also provide evaluation and rehabilitation since, for people younger than 18, hearing loss may cause problems in language development and educational and social growth. An audiologist is qualified by training and experience to assist in the evaluation and rehabilitation of hearing loss in people vounger than 18.

Table of Contents

| Hearing aid information |
|---|
| FDA warnings and cautions (US only) |
| Introduction |
| Getting to know your hearing aid |
| Preparing your hearing aids for use |
| Placing your hearing aids in your ears |
| Removing your hearing aids from your ears |
| Using your hearing aids |
| Advanced options |
| Cleaning and caring for your hearing aids |
| Wireless accessories 55 |
| Tinnitus Management 5- |
| General warnings and cautions 65 |
| Troubleshooting |
| Regulatory information 79 |
| Technical specifications 7 |
| Additional information |

Introduction

Thank you for choosing ReSound hearing aids. We recommend that you use your hearing aids every day.



NOTE: Read this booklet carefully BEFORE using your hearing aids.

Intended use

Generic air-conduction hearing aids are wearable sound-amplifying devices intended to compensate for impaired hearing. The fundamental operating principle of hearing aids is to receive, amplify, and transfer sound to the eardrum of a hearing impaired person.

The hearing aid is intended to be used by adults and children 12 years of age or older. The hearing aid is intended to be used by lay persons.


For devices including a dome

The dome is intended to be connected to a receiver tube on the hearing aid. The dome is intended to ensure that the sound outlet of the hearing aid is placed in the ear canal.

This accessory is intended to be used by the same age group as the hearing aid. The accessory is intended to be used by lay persons.


Getting to know your hearing aid

Rechargeable model



^{*}Open dome is shown. Your own may look different.


Zinc-Air battery model (312)



^{*}Open dome is shown. Your own may look different.

^{**}Inside battery compartment.

Zinc-Air battery model (13)



^{*}Open dome is shown. Your own may look different.

^{**}Inside battery compartment.

Domes and earmolds

Domes are all light grey.











Tulip dome

Closed dome

Open dome

Power dome

Custom earmold

Receivers and sport locks



Microphone



Left / right

colour coding

Your receiver may

shown here

have a microphone as

Your hearing aids are color-coded

Left = blue. Right = red.

Dome Sport lock

The sport lock helps keep your hearing aids in place when you are physically active.

Preparing your hearing aids for use

Battery warnings - all models



WARNING: Batteries are harmful for the environment. Therefore, never try to burn them. Dispose of your used batteries according to your country's regulations or return them to your hearing care professional.

Battery warnings - Rechargeable model



WARNING:

- Rechargeable hearing aids are supported by a non-removable rechargeable lithium-ion battery cell. Exposure or ingestion can be seriously harmful.
- Never put your rechargeable hearing aid in your mouth.
- Keep your rechargeable hearing aid away from pets, children, and people with cognitive, intellectual, or mental health challenges.
- Never swallow lithium-ion batteries nor place them inside any part of the body, as this may lead to fatal injuries in two hours or less. If a lithium-ion battery is swallowed or placed inside any part of the body. or if a rechargeable hearing aid is swallowed, seek immediate medical attention.
- If the outer casing of your rechargeable hearing aid is broken, the rechargeable batteries inside may leak. In this case, do not attempt to use the hearing aid and contact your hearing care professional.

 Battery leakage can cause chemical burns. If you get exposed to battery leakage material, rinse immediately with warm water. If you get chemical burns, redness, or skin irritation from battery leakage, seek medical attention.

Battery warnings - Zinc-air battery models



WARNING:

- Never put a hearing aid battery in your mouth.
- Keep hearing aid batteries away from pets, children, and people with cognitive, intellectual, or mental health challenges.
- Never swallow batteries nor place them inside any part of the body, as the battery can cause serious
 injuries. If a battery has been swallowed or placed inside any part of the body, seek immediate medical
 attention.




CAUTION:

- Batteries may leak. If you are not going to use your hearing aids for a few days, you MUST remove the batteries.
- Battery leakage can cause chemical burns. If you are exposed to a battery leakage, rinse immediately
 with lukewarm water. If you get chemical burns, redness, or skin irritation, seek immediate medical
 attention.
- Never try to charge zinc-air batteries as this can cause leakage or a small explosion.



NOTE: Always use new zinc-air batteries that have a minimum remaining shelf life of one year.

Inserting the battery into your hearing aid — Zinc-air battery models









- 1. Open the battery door with vour fingernail. Remove the old battery (if there is one).
- 2. Remove the packaging and protective foil from the new battery. Wait for two minutes while the battery activates
- 3. Insert the new battery, with the plus (+) sign correctly positioned, into the battery compartment door. Do not insert it directly into the hearing aid.
- 4. Close the battery door.



- To save battery power, turn off your hearing aids when they are not in use.
- At night, turn off the hearing aids. Open the battery door completely to allow moisture to evaporate and prolong the hearing aids' life span.

If the hearing aids are experiencing frequent loss of connection to wireless accessories, contact your hearing care professional for a list of low impedance batteries.

Charging your hearing aids — Rechargeable model

We recommend that you fully charge your hearing aids before using them. To learn how to charge your hearing aids, please consult the user quide for your hearing aid charger.

Low battery warning (All models)

When the batteries are low on power, your hearing aids' volume goes down, and they play a melody every 15 minutes, until they are empty and turn off.



NOTE: Keep extra batteries with you. (Zinc-air battery models)

Low battery alert when paired with wireless accessories (all models)



NOTE:

Your batteries will run out faster when you use wireless functions like streaming from your smart device or from your TV with our TV streamer. As the battery power declines, the wireless functions stop working, A short melody will play every five minutes to let you know that the battery power is low. The table below shows the functionality with different battery charge levels.

| Battery level | Signal | Hearing aid | Remote control | Streaming |
|---------------|-----------------------------------|-------------|----------------|-----------|
| Fully charged | | ✓ | ✓ | ✓ |
| Low | 4 even tones | √ | √ | х |
| Depleted | 3 even tones and 1 longer tone | √ | х | х |

These will work again when you insert a new battery (Zinc-air battery models) or charge the hearing aid (Rechargeable model).

Placing your hearing aids in your ears

How to tell left from right

Left hearing aid (blue marking on the receiver)

Right hearing aid (red marking on the receiver)








CAUTION: If you have two hearing aids, they may each be programmed differently. Do not swap them as this could damage your hearing.

Your hearing aids are color-coded. Left = blue. Right = red.

If your hearing aids are not color coded, ask your hearing care professional to add color coding.

Inserting the receiver dome into your ear

If your hearing aids have domes, follow these instructions:



- 1. Hang the hearing aid over the top of your ear.
- 2 Hold the receiver wire where it bends and gently place/push the receiver dome into your ear canal
- 3. Push the dome far enough into your ear canal so that the wire rests against your head. You can check in a mirror



NOTE: To avoid whistling, ensure that the wire, and dome fit correctly into your ear. If you continue to experience whistling, check the troubleshooting guide for other possible reasons and solutions.

Sport lock

If you are very active, your hearing aids may move from the correct position. To avoid this situation, your hearing care professional can attach and adjust a sports lock to the receiver.

To insert a hearing aid with a sports lock:

- 1. Insert the hearing aid as usual
- 2. Tuck the sports lock into the bottom of the concha (the part of your ear just outside the ear canal).






NOTE: Sports locks may become stiff, brittle, or discolored over time. Contact your hearing care professional for a replacement. Your hearing care professional will help you replace it.

Inserting earmolds into your ears


If your hearing aids have earmolds, follow these instructions:



1 Hold the earmold between your thumb and index finger and position the sound outlet into vour ear canal.



2. Slide the earmold into your ear with a 3. gentle, twisting movement. Move the earmold up and down and press gently. Opening and closing your mouth may be helpful.



Place the hearing aid behind your ear and ensure it is secure. When properly inserted, your hearing aids should fit snugly and comfortably.




NOTE: When inserting an earmold, it may help to pull your ear up and back with the opposite hand.


Removing your hearing aids from your ears

Removing receiver domes from your ears

If your hearing aids have domes, follow these instructions:




1. Lift the hearing aid off your ear.




2. Hold the receiver wire with your thumb and forefinger where it bends and pull the dome out of your ear canal.

Removing earmolds from your ears

If your hearing aids have earmolds, follow these instructions:





- 1. Lift the hearing aid from behind your ear. Let it hang beside your ear momentarily.
- 2. Using your thumb and index finger, gently pull the earmold (not the hearing aid or the wire) loose from your ear. If your earmold has a removal cord, pull it gently to assist. The removal cord is a separate line that is additionally attached to the earmold if requested. Remove the earmold completely by twisting it gently.

Using your hearing aids

Turning your hearing aids on and off

You can turn on your hearing aids either before or after you place them in your ears. Your hearing aids always start on program 1 at the pre-set volume.

Turning your hearing aids on and off — Rechargeable model

To turn your hearing aids on or off, press the push button for **5 seconds**.

Pressing the button for different lengths of time enables other functions such as flight mode and streaming. These are described elsewhere in this manual.



Explanation of indicator lights:

- One 2-second green blink: The hearing aid turns on and is in operating mode.
- Three 1-second green blinks: The hearing aid turns off.
- When the hearing aid is present in the charger, green indicator lights pulsate: The hearing aid is recharging.

Your hearing aids turn off automatically when you place them in the charger and they turn on automatically when you remove them from the charger.

Turning your hearing aids on and off — Zinc-air battery models



To turn your hearing aid on, close the battery door.

To turn the hearing aid off, open the battery door (with your fingernail).


Smart Start


Smart Start delays the time that your hearing aid turns on after you close the battery door (Zinc-Air battery models) or remove the hearing aids from the charger (Rechargeable model). This function produces a beep for each second of the delay period (either 5 or 10 seconds). Smart Start delays when sound comes through your hearing aid. It gives you time to put on the hearing aid without whistling or other issues.

If you do not want to use this function, ask your hearing care professional to deactivate it.

Using the push button/multi-function button

These buttons allow you to use a variety of listening programs. Each program is suitable for a different situation.





Multi-function button

Listening programs

Your hearing care professional can activate one or more listening programs in your hearing aids. These programs can help you in specific situations. Ask your hearing care professional about which programs could be useful for you.

| Programs | Use |
|----------------------|---|
| All-Around | The best option if you want only one program. |
| Ultra Focus | For use in very noisy places (more focused than the Restaurant program). |
| Restaurant | For use in noisy places such as restaurants and social gatherings. |
| Music | For listening to music. |
| Acoustic phone | A special program for phone conversations. |
| Outdoor | For outdoor use (to reduce wind noise). |
| Telecoil phone + Mic | For use if you have a phone with a telecoil. |
| Telecoil loop + Mic | For use in places that have a teleloop system such as theaters and places of worship. |

Apps

We have an app that you can use to control your hearing aids. You can use the app to adjust the volume, change programs and stream from another device. There is more information on these in the Advanced Options chapter.

Using the push button

The "312" models have a push button.

Changing programs

- Push the button to switch from one program to the next. Your hearing aids will beep to confirm the change. The number of beeps indicates which program you have selected.
- To get back to the default listening program, keep scrolling, or turn your hearing aids off and on again.



NOTE: If you have two hearing aids with synchronization enabled, changing the program on one hearing aid automatically applies the change to the second hearing aid as well.

Streaming sound to your hearing aids

Press the button and keep it pressed for three seconds. Your hearing aids will start streaming from the audio device you have chosen

Before you can stream to your hearing aids, you must pair them with your audio device. For more information on how to do this, consult the user's instructions for your audio device.



Using the multi-function button



The "13" models have a multi-function button

It is called a multi-function button because it has several functions; you can change program, control volume, and activate streaming.

Changing programs

- To change the program, press the top of the multifunction button and hold it for three seconds. Your hearing aids will beep to confirm the change. The number of beeps indicates which program you have selected
- Continue pressing it to scroll through the programs available to you.
- To get back to the default listening program, keep scrolling, or turn your hearing aids off and on again.





NOTE: If you have two hearing aids with synchronization enabled, changing the program on one hearing aid automatically applies the change to the second hearing aid as well.


This feature can also be configured to allow you to use one hearing aid to increase volume and the other to decrease volume. Any increase or decrease made on one hearing aid will also apply to the other to ensure the volume is the same in both hearing aids.

Your hearing care professional can change the default settings and fill in the table below to record the new settings:

| Button action | Default setting | New setting |
|-----------------------------|--------------------|-------------|
| Short press up | Increase volume | |
| Short press down | Decrease volume | |
| Long press up (3 seconds) | Change program | |
| Long press down (3 seconds) | Activate streaming | |

Controlling the volume

If your hearing care professional has activated your volume control, follow these instructions:



To increase the volume, briefly press the top part of the button



To reduce the volume, briefly press the **bottom** part of the button.

Streaming sound to your hearing aids

Press the bottom part of the button and keep it pressed for three seconds. Your hearing aids will start streaming from the audio device you have chosen.

Before you can stream to your hearing aids, you must pair them with your audio device. For more information on how to do this, consult the user's instructions for vour audio device.



Telecoil (Optional)

Your hearing aid may have a telecoil. The Telecoil function may help to improve understanding of speech with Hearing Aid Compatible (HAC) telephones and in theaters, movie theaters, houses of worship, etc. that have a teleloop installed.

When you select the Telecoil function, your hearing aid picks up signals from the teleloop or your HAC telephone. Your hearing care professional can activate the Telecoil program.



NOTE:

- The telecoil does not work without a teleloop (i.e., an induction loop) or an HAC telephone.
- If you are having trouble hearing with the telecoil, ask your hearing care professional to modify the function
- If there is no sound from your hearing aids in a teleloop system with an active Telecoil function, the teleloop system may not be turned on or may not be operating correctly.
- The sound from the teleloop and the hearing aid microphones can be mixed according to your preference. Ask your hearing care professional for more details.

Advanced options

Using a telephone

Your hearing aid allows you to use your telephone as you normally do. Finding the optimal position for holding the phone may require practice.

The following suggestions may be helpful:

- 1. Hold the telephone up to your ear canal or hold it close to the hearing aid microphones as illustrated.
- 2. If you hear whistling, try holding the telephone in the same position for a few seconds. The hearing aid may be able to cancel the whistling.
- 3. You can also try holding the telephone slightly away from the ear.



Depending on your needs, your hearing care professional may activate a function specifically for telephone use.

Mobile phones

Your hearing aids comply with the most stringent Standards of International Electromagnetic Compatibility. Any degree of disturbance can be due to the nature of your particular mobile phone or of your wireless telephone service provider.



Using your hearing aids with iPhone, iPad, and iPod touch (optional)

Your hearing aids are Made for iPhone®, iPad®, and iPod touch®, which allow for direct audio streaming, including hands-free phone calls from the latest iPhone and iPad models¹, and control from these mobile devices.

Streaming from an Android™ smartphone

Some Android smartphones can stream audio, including phone calls, directly to your hearing aids. Your device must be running Android 10 or newer and it must have the Android Streaming for Hearing Aids feature as well.



NOTE: For assistance with pairing and using these products with your hearing aids, contact your hearing care professional.

Using your hearing aid with smartphone apps (optional)

The smartphone apps send and receive signals to and from the hearing aids via smartphones.

- Do not disable app notifications.
- Install updates to keep the app working correctly.
- To get a printed version of the smartphone app user's guide, please refer to information on the back page of this user's guide or consult customer support.

¹Hands-free calls are compatible with iPhone 11 or later, iPad Pro 12.9-inch (5th generation), iPad Pro 11-inch (3rd generation), iPad Air (4th generation), and iPad mini (6th generation), with software updates iOS 15.3 and iPadOS 15.3 or later.



- For assistance with pairing and using these products with your hearing aid, contact your hearing care professional or visit our support site...
- If your Bluetooth® enabled Android smartphone does not stream directly to your hearing aids, you can use our ReSound Phone Clip+ for streaming capabilities.

ReSound Assist (optional)

If you have signed up to use ReSound Assist with your hearing aids, you can have your hearing aids adjusted remotely without having to visit your hearing care professional. All you need is a smart device that is Internet enabled. This allows you to experience freedom and flexibility such as:

- 1. Requesting remote assistance to adjust your hearing aids.
- 2. Ensuring the best possible performance of your hearing aids by keeping them up to date with the latest software



NOTE: Your hearing aids will turn off during the install and update process.

For optimum performance, ensure your hearing aids are connected to the ReSound Smart 3D™ app and placed close to the iPhone, iPad, iPod touch, or Android smartphone.

This service only works if your mobile device is connected to the internet. Your hearing care professional will provide information about this option, and how it works with the ReSound Smart 3D™ app.





Phone Now (optional)

If you place a magnet on the telephone receiver, your hearing aids will automatically switch the telephone program on when the receiver is close to your ear.

When you remove the receiver from your ear, the hearing aids automatically return to the previous listening program.



NOTE: Ask your hearing care professional to enable Phone Now as one of your programs.

Phone Now warnings




WARNING:


- Swallowing a magnet can be harmful to your health. If a magnet is swallowed, seek immediate medical attention
- · Keep magnets out of reach of pets, children, and people with cognitive, intellectual, or mental health challenges.
- The Phone Now magnet may affect sensitive medical devices/electronic systems. Seek advice from the manufacturers regarding appropriate safety measures when using the Phone Now solution near the sensitive device/equipment (pacemakers and defibrillators) in question. If the manufacturer cannot issue a statement, we recommend keeping the magnet or a telephone equipped with the magnet 30 cm (12") away from magnetically sensitive devices (e.g., pacemakers).

Placing the Phone Now magnet

Place the magnet on your telephone receiver as follows:







2 Remove the foil from the magnet.



3. Place the magnet on the phone.



CAUTION:

- If you experience frequent signal loss or noise during calls, move the Phone Now magnet to another place on the telephone receiver.
- Only use magnets supplied by ReSound.

How to use Phone Now

- 1. Lift the telephone to your ear.
- 2. When you hear a short melody, the phone program is active.



- You may need to move the telephone receiver slightly to find the best position for reliable Phone Now activation and a good hearing experience on the telephone.
- If your hearing aids have the Comfort Phone functionality enabled, the hearing aid on the non-phone ear automatically attenuates.
- Do not cover the phone loudspeaker opening with the magnet.
- If the function does not work to your satisfaction, moving the magnet to another position may improve ease of use and comfort
- If your hearing aids do not switch to the telephone program consistently, try repositioning the magnet or adding additional magnets.

Flight Mode (optional)

Your hearing aids can be controlled from your smartphone or Remote Control, an option that can be added by your hearing care professional. However, in some areas you are requested to turn off wireless communication.



CAUTION:

 This hearing aid contains a radio frequency transmitter. When boarding a flight, follow airline instructions and turn off hearing aid wireless functionality when and as directed.



NOTE: You must follow the processes below for both hearing aids, even if synchronization is enabled.

Turning off wireless communication (activating Flight Mode) - Zinc-air battery models

- 1. Open and close the battery door on each hearing aid three times within 10 seconds.
- 2. A 10-second double tone () means the hearing aid is now in Flight Mode.

It is important to wait an additional 15 seconds after you activate Flight Mode before opening and closing the battery compartment again for any reason. If you open and close the battery compartment during this 15 second window, wireless communication will reactivate.

Activating wireless communication (turning off Flight Mode) - Zinc-air battery models

- 1. Open and close the battery door on each hearing aid once.
- 2 Wireless communication will be activated after 10 seconds

Turning off wireless communication (activating Flight Mode) - Rechargeable model

- 1. Turn off your hearing aid.
- 2 Press the button for 9 seconds
- 3. Your hearing aid will double-flash four times. If you are wearing your hearing aids at the time, you will hear double tones (,, ,) for about 10 seconds, meaning the hearing aid is now in Flight Mode.

Activating wireless communication (turning off Flight Mode) - Rechargeable model

- 1. Turn your hearing aid off and then on.
- 2 Wireless communication will be activated after 10 seconds

Cleaning and caring for your hearing aids

Cleaning tools

These cleaning tools come with your hearing aids:



- 1. Soft cloth.
- 2. A brush with a battery magnet.



Your hearing care professional may give you a set of wax filters

General instructions for care and maintenance



NOTE: To ensure you get the highest quality experience and longest useful lifetime out of your hearing aids, it is important to clean and care for them.

Keeping your hearing aids in perfect working order is easily done. Just follow these steps:

- 1. When you remove your hearing aids, turn them off by opening the battery doors completely. This helps to dry them out. (Zinc-air battery models.)
- 2. After removing your hearing aids, wipe them with a soft cloth to keep them clean and dry.
- 3. If you use a drying agent, only use recommended products.
- 4. Apply cosmetics, perfume, after-shave, hairspray, lotions etc. BEFORE putting on your hearing aids. These products can damage or discolor your hearing aids.
- 5. Never immerse your hearing aids in liquid.
- 6. Keep your hearing aids away from excessive heat and direct sunlight.
- 7. Do not swim, shower, or take steam baths while wearing your hearing aids.



CAUTION:

- Never use alcohol or other cleaning solutions to clean your hearing aids. This can damage your hearing aids and may cause a skin reaction.
- Ear wax or other residue on your hearing aids can cause an infection. To avoid this, clean your hearing aids as instructed




Daily care and maintenance

It is important to keep your hearing aids clean and dry on a daily basis. Use the supplied cleaning tools.



1. Wipe your hearing aids with the cloth



2. Swipe the small brush across the microphones. If your receiver has a microphone, remember to brush that one as well.



- · Never try to put the bristles of the small brush or the cleaning wire into the microphone inlets. This can damage your hearing aids.
- Do not use water to clean your receiver wires and domes or earmolds.
- Use a soft, dry cloth to wipe your receivers and domes or earmolds clean.
- Use a soft, dry cloth to wipe your earmolds clean.

Receiver wires may become stiff, brittle or discolored over time. If you think your receiver wire needs to be replaced, contact your hearing care professional.

Changing your hearing aid domes

We recommend that your hearing care professional shows you how to change domes. You should change your domes every 3 months or more often according to the advice of your hearing care professional.



NOTE: Incorrect dome replacement can result in a dome being left in your ear when you remove your hearing aid.



CAUTION:

If you suspect that you have a dome stuck in your ear, consult your hearing care professional.

Standard domes

The illustrations show an open dome, but the procedure is the same for all other domes. Follow these instructions to replace your domes.



1. Remove the used dome by pulling it off the receiver and then discard it. This may require a bit of force.



3. Check that the dome is securely mounted by carefully lifting the lower part of the dome and verifying that the collar completely covers the ribbed end of the receiver.




2 Push the new dome over the ribbed end of the receiver




4. When placed correctly, the dome will appear slightly angled.

Tulip domes


To change a tulip dome, follow these instructions:







- 1. Remove the used dome by pulling it off the receiver and then discard it. This may require a bit of force.
- 2. Push the largest petal back, then press the tulip dome over the ribbed end of the receiver





3. Check that the tulip dome is securely 4. Push the larger tab forward. mounted by verifying that the collar completely covers the ribbed end of the receiver






This illustration shows a correctly mounted tulip dome. Ensure the large petal is outside the small petal.

Changing the wax filter

The wax filter helps keep earwax away from the components of the hearing aid. You must replace it on a regular basis. Consult your hearing care professional for advice on how often you need to change them. It will depend on how much wax your ears produce.

If you wear a dome, remove it before following this process for replacing the wax filter. See "Changing your hearing aid domes", page 44.

You will need your box of wax filter tools.




Box of 8 wax filter tools.


The wax filter tool has two functions: a removal tip to collect the used filter, and a replacement tip with a white filter.


Removing the old wax filter






- 1 Remove the dome from the receiver
- 2. Open the wax filter case and take out one of the tools. Each tool has a small hook in one end and a new wax filter in the other.






3. Insert the removal tip into the used wax filter and then pull the tool straight out. It is important to pull it straight and not on an angle.


Inserting the new wax filter



1 Insert the other end of the tool into the sound outlet (the end with the replacement filter).



2. Gently press the replacement straight into the sound outlet until the outer ring is touching the sound outlet



3. Pull the tool straight out. Your new wax filter will remain in place. Re-attach your dome or a replacement dome.

Storing your hearing aids — Zinc-air battery models



1. The best storage for your hearing aids is the case they came in.

2. When they are not being worn. keep your hearing aids dry by leaving the battery doors open.





WARNING:

- Never put a hearing aid battery in your mouth.
- Keep hearing aid batteries away from pets, children, and people with cognitive, intellectual, or mental health challenges.
- Never swallow batteries nor place them inside any part of the body, as the battery can cause serious injuries. If a battery has been swallowed or placed inside any part of the body, seek immediate medical attention

Wireless accessories

Our wireless eco-system features a comprehensive range of seamlessly integrated wireless accessories. These let you stream high quality stereo sound and speech directly to your hearing aids.

Please find the list of available wireless accessories and their features below:

- ReSound TV Streamer 2 streams audio from a TV and most other audio sources to your hearing aids at a volume that suits vou.
- **Remote Control** adjusts volume, mutes your hearing aids, and changes programs.
- ReSound Remote Control 2 adjusts volume, mutes your hearing aids, changes programs, and displays vour settings.
- ReSound Phone Clip+ streams phone conversations and stereo sound to both hearing aids and doubles as a remote control
- ReSound Micro Mic is a microphone that can be worn by others. It improves speech comprehension in noisy situations
- ReSound Multi Mic works like the ReSound Micro Mic but doubles as a table microphone. It connects with loop and FM systems and has a mini-jack input for streaming audio from a computer or music player.



- · Ask your hearing care professional for more information on the range of ReSound wireless accessories
- You should only use ReSound wireless accessories with your wireless hearing aids. For assistance with pairing your hearing aid with any of ReSound accessories, please refer to the user's guide for the relevant wireless accessory.



Tinnitus Management

Tinnitus Sound Generator module

Your ReSound hearing aid includes the Tinnitus Sound Generator (TSG) module, a tool for generating sounds to be used in tinnitus management programs to temporarily relieve suffering from tinnitus. The TSG can generate sounds adjusted to the specific therapeutic needs and your personal preference as determined by your doctor, audiologist, or hearing care professional. Depending on the selected hearing aid program and the environment you are in, you will sometimes hear the therapeutic sound resembling a continuous or fluctuating noise.

Indications for use of the TSG module

The Tinnitus Sound Generator module is a tool to generate sounds to be used in a Tinnitus Management Program to temporarily relieve patients suffering from Tinnitus. The target population is primarily the adult population over 18 years of age. This product may also be used for children 12 years of age or older. However, children and physically or mentally disabled users will require training by a doctor, audiologist, hearing care professional, or their guardian in the insertion and removal of the hearing aid containing the TSG module.

For healthcare professionals

The Tinnitus Sound Generator module is targeted for healthcare professionals who are treating patients suffering from Tinnitus, as well as conventional hearing disorders. The initial fitting of the Tinnitus Sound Generator module must be done during an in-office visit by a hearing professional participating in a Tinnitus Management Program, If deemed feasible by the hearing professional, subsequent fittings of the Tinnitus Sound Generator

module may be performed remotely and in real time while having live communication via live audio, video, and chat on the user's dedicated app.

User instructions for the TSG module

Description of the device

The Tinnitus Sound Generator (TSG) Module is a software tool that generates sounds to be used in tinnitus management programs to temporarily relieve suffering from tinnitus.

Explanation of how the device functions

The TSG module is a frequency and amplitude shaped white-noise generator. The noise signal level and frequency characteristics can be adjusted to the specific therapeutic needs as determined by your doctor. audiologist, or hearing care professional.

Your doctor, audiologist, or hearing care professional can modulate the generated noise in order to make it more pleasant. The noise can then resemble, for example, breaking waves on a shore.

Modulation level and speed can also be configured to your likes and needs. An additional feature can be enabled by your hearing care professional that allows you to select predefined sounds that simulate sounds from nature, such as breaking waves or running water.

If you have two wireless hearing aids that support ear-to-ear synchronization, this functionality can be enabled by your hearing care professional. This will cause the Tinnitus Sound Generator to synchronize the sound in both hearing aids.

If your tinnitus only troubles you in quiet environments, your doctor, audiologist, or hearing care professional can set the TSG Module so that it becomes audible exclusively in such surroundings. The overall sound level can be

adjusted using an optional volume control. Your doctor, audiologist, or hearing care professional will review with you the need for having such a control.

For hearing aids where ear-to-ear synchronization is enabled, your hearing care professional can also enable environmental monitoring synchronization so that the TSG noise level is automatically adjusted simultaneously in both hearing aids dependent on the background sound level. Additionally, since the hearing aid has a volume control, the background noise level monitored by the hearing aid and the volume control can be used simultaneously to adjust the generated noise level in both hearing aids.

The scientific concepts that form the basis for the device

The TSG module provides sound enrichment with the aim of surrounding the tinnitus sound with a neutral sound which is easily ignored. Sound enrichment is an important component of most approaches to tinnitus management, such as tinnitus retraining therapy (TRT).

To assist habituation to tinnitus, this needs to be audible. The ideal level of the TSG module, therefore, should be set so that it starts to blend with the tinnitus, and so that you can hear both your tinnitus as well as the sound used

In a majority of instances, the TSG module can also be set to mask the tinnitus sound, so to provide temporary relief by introducing a more pleasant and controllable sound source.

TSG volume control

The sound generator is set to a specific loudness level by the hearing care professional. When switching the sound generator on, the volume will have this optimal setting. Therefore, it might not be necessary to control the volume (loudness) manually. However, the volume control provides the ability to adjust the volume, or amount of stimulus, to the liking of the user. The tinnitus sound generator volume can only be adjusted within the range set by the hearing care professional.

The volume control is an optional feature in the TSG module used for adjusting the sound generator output level.

Using TSG with smartphone apps

The tinnitus sound generator control via hearing aid push buttons can be enhanced with wireless control from a TSG control app on a smartphone or mobile device. This functionality is available in supported hearing aids when a hearing care professional has enabled the TSG functionality during fitting of the hearing aid.



NOTE: To use smartphone apps, the hearing aid must be connected with the smartphone or mobile device

TSG - Technical specifications

Audio signal technology: Digital.

Available sounds

White noise signal which can be shaped with the following configurations:

| High-pass filter | Low-pass filter |
|------------------|-----------------|
| 500 Hz | 2000 Hz |
| 750 Hz | 3000 Hz |
| 1000 Hz | 4000 Hz |
| 1500 Hz | 5000 Hz |
| 2000 Hz | 6000 Hz |
| - | 8000 Hz |

The white noise signal can be modulated in amplitude with an attenuation depth of up to 14 dB.



Prescription use of this device

The TSG module should be used as prescribed by your doctor, audiologist or hearing healthcare professional. In order to avoid permanent hearing damage, the maximum daily usage depends on the level of the generated sound.

To adjust TSG, please consult your hearing healthcare professional.

Should you develop any side effects from using the sound generator, such as dizziness, nausea, headaches, perceived decrease in auditory function, or increase in tinnitus perception, you should discontinue use of the sound generator and seek medical evaluation.

Target population

The target population is primarily the adult population over 18 years of age. This product may also be used for children 12 years of age or older. However, children and physically or mentally disabled users will require training by a doctor, audiologist, hearing care professional, or their guardian in the insertion and removal of the hearing aid containing the TSG module.

Important notice for prospective sound generator users

A tinnitus masker is an electronic device intended to generate noise of sufficient intensity and bandwidth to mask internal noises. It is also used as an aid in hearing external noises and speech.

Good health practice requires that a person with a tinnitus condition have a medical evaluation by a licensed physician (preferably a physician who specializes in diseases of the ear) before using a sound generator. Licensed physicians who specialize in diseases of the ear are often referred to as otolaryngologists, otologists, or otorhinolaryngologists.

The purpose of medical evaluation is to assure that all medically treatable conditions that may affect tinnitus are identified and treated before the sound generator instrument is used.

The sound generator instrument is a tool to generate sounds to be used with appropriate counseling and/or in a tinnitus management program to relieve patients suffering from tinnitus.

Warning information



- Sound generators can be dangerous if improperly used.
- Sound generators should be used only as advised by your doctor, audiologist, or hearing care professional.
- Sound generators are not toys and should be kept out of reach of anyone who might cause themselves injury (especially children and pets).



CAUTION:

- Should the user develop any side effects from using the sound generator, such as dizziness, nausea. headaches, a perceived decrease in auditory function or an increase in tinnitus perception, the user should discontinue use of the sound generator and seek medical evaluation.
- To prevent unintended usage by pediatric, or physically or mentally disabled users, the volume control must, if enabled, be configured to only provide a decrease of the sound generator output level.
- Children, and physically or mentally disabled users will require guardian supervision while wearing the TSG hearing aid.



Tinnitus Sound Generator warning to hearing care professionals

A hearing care professional should advise a prospective sound generator user to consult promptly with a licensed physician (preferably an ear specialist) before getting a sound generator, if the hearing care professional determines through inquiry, actual observation, or review of any other available information concerning the prospective user that the prospective user has any of the following conditions:

- 1. Visible congenital or traumatic deformity of the ear.
- 2. History of active drainage from the ear within the previous 90 days.
- 3. History of sudden or rapidly progressive hearing loss within the previous 90 days.
- 4 Acute or chronic dizziness
- 5. Unilateral hearing loss of sudden or recent onset within the previous 90 days.
- 6. Audiometric air-bone gap equal to or greater than 15 dB at 500 Hertz (Hz), 1000 Hz, and 2000 Hz.
- 7. Visible evidence of significant cerumen accumulation or a foreign body in the ear canal.
- 8 Pain or discomfort in the ear



CAUTION: The maximum output of the sound generator falls into the range that can cause hearing loss according to OSHA regulations. In compliance with NIOSH recommendations, the user should not use the sound generator for more than eight (8) hours a day when it is set to a level of 85 dB SPL or above. When the sound generator is set to levels of 90 dB SPL or above, the user should not use it for more than two (2) hours per day. In no case should the sound generator be worn at uncomfortable levels.

General warnings and cautions



WARNING:

- · Consult a hearing care professional:
 - If you think there may be a foreign object in your ear canal
 - If you experience skin irritation
 - If excessive ear wax accumulates with the use of the hearing aid
- Do not wear your hearing aids while being exposed to radiation. Some types of radiation, such as from MRI or CT scanners, can affect the settings in your hearing aids, causing malfunction and potentially damage to your hearing.
- Other types of radiation, such as burglar alarms, room surveillance systems, mobile phones, and radio
 equipment will not damage your hearing aids. However, they may briefly affect the sound quality in
 your hearing aids and may create undesired sounds.
- Never use your hearing aids in places with explosive gases such as mines, oil fields, or similar locations unless these areas are certified for hearing aid use. Using your hearing aids in places that are not certified for hearing aid use can be dangerous.
- For safety reasons, use only the charger supplied by ReSound. (Rechargeable model).
- Do not attempt to dry your hearing aids in an oven, microwave oven, or other heating equipment. This
 will cause them to melt and may cause burns to your skin.

 External devices connected to the electrical input must be safe according to the requirements of IEC 60601-1, IEC 60065, EN/IEC 62368-1, or IEC 60950-1, as appropriate (wired connection, for example HI-PRO. SpeedLink).

Warnings related to power hearing aids

- A power hearing aid can produce very loud sound to compensate for a severe or profound hearing loss. Therefore, there is risk of further impairing the remaining hearing.
- Your hearing aids have been customized to amplify soft and loud sounds according to your particular needs. If the amplification seems too loud or you suspect the hearing aid is malfunctioning (e.g., you hear distorted or unusual sound), contact your hearing care professional. A malfunctioning hearing aid can damage your hearing.
- In general, exposure to loud sounds can damage your hearing. This could be loud music or loud environments. You can best protect your hearing by reducing exposure to loud sound and environments or by using hearing protection.



CAUTION:

- Use your hearing aids as your hearing care professional recommends. Incorrect use may damage vour hearing.
- Do not use a broken or modified hearing aid. It may not work properly and may be harmful to your hearing. It may also cause scratches or sores due to sharp edges.
- Use only original consumables from ReSound(e.g., domes).
- Use only accessories intended for use with your hearing aids. Consult your hearing aid professional for more information

- Do not try to modify the shape of your hearing aid or accessories. This can cause skin reactions or sharp edges leading to scratches or sores.
- If you have two hearing aids, they may each be programmed differently. Do not swap them as this could damage your hearing. Your hearing aids are color-coded. Left = blue. Right = red. If your hearing aids are not color coded, ask your hearing care professional to add color coding.
- If you experience side effects, contact a hearing care professional. Possible side effects from wearing a hearing aid may be:
 - Dizziness
 - Tinnitus
 - Perceived worsening of hearing loss
 - Nausea
 - Headache
 - Skin reaction
 - Far wax accumulation
- If you suspect that you have a dome, wax filter, or other object in your ear canal, consult your hearing care professional. These objects can be harmful and can cause an infection in your ear.
- If you have a sore or injury where your hearing aid touches your ear or head, continued use of the hearing aid may cause it to worsen or prevent it from healing. Consult a hearing care professional for assistance
- Your hearing aids are tuned to your hearing. Do not allow others to use your hearing aids as this can damage their hearing.

- When using wireless functions, your hearing aid uses low-powered digitally coded transmissions to communicate with other wireless devices. It is possible, but not likely, that other electronic devices will be affected. If this happens, move the hearing aid away from the affected electronic device.
- Turn off hearing aid wireless functionality when this is required.

For hearing care professionals



CAUTION: Do not change the outer casing or any parts of a hearing aid unless appropriately protected against ESD.

Troubleshooting

| Issue | Potential cause | Potential solution |
|-------------------------|---|--|
| Feedback or "whistling" | Is your earmold or dome inserted correctly in your ear? | Put it back in. |
| | Is the volume very loud? | Reduce the volume. |
| | Is the receiver wire broken or the earmold clogged? | Visit your hearing care professional. |
| | Are you holding an object (e.g., a hat or a telephone mouthpiece) close to the hearing aid? | Move your hand away to create more space between the hearing aid and the object. |
| | Is your ear full of wax? | Visit your doctor. |

| Issue | Potential cause | Potential solution |
|----------|--|--|
| No sound | Is the hearing aid turned on? | Turn it on. |
| | Is there a battery in the hearing aid? (Zinc-air battery models) | Insert a new battery. |
| | Is the battery still good? (Zinc-air battery models) | Replace with a new battery. |
| | Is the hearing aid charged? (Rechargeable model) | Charge the hearing aid. (See the user manual for your charger.) |
| | Is the hearing aid charger charged? (Rechargeable model) | Charge the hearing aid charger. (See the user's guide for your charger.) |
| | Is the receiver wire broken or the earmold clogged? | Visit your hearing care professional. |
| | Is your ear full of wax? | Visit your doctor. |

| Issue | Potential cause | Potential solution |
|---|---|--|
| Sound is distorted, spluttering, or weak? | Is the battery dead? (Zinc-air battery models) | Replace with a new battery. |
| | Is the battery dirty? (Zinc-air battery models) | Clean it or replace it with a new one. |
| | Is the receiver wire broken or the earmold clogged? | Consult your hearing care professional. |
| | Is there moisture in your hearing aid? | Use a desiccant (drying kit). |
| Battery drains very quickly. | Did you leave your hearing aid on for long periods of time? | Always turn off your hearing aids when you are not using them, or place them in the charger for charging (Rechargeable hearing aids). |
| | | Your batteries will run out faster when you use wireless functions like streaming from your smart device or from your TV with our TV streamer. |
| | Is the battery old? (Zinc-air battery models) | Check the battery packaging. |
| | Is the hearing aid 3-4 years old? (Rechargeable model) | Visit your hearing care professional. |

| Issue | Potential cause | Potential solution |
|-----------------------------|--|---|
| Hearing aid is not charging | Is the hearing aid correctly inserted in the charger? | Put the hearing aid back in the charger. (See the user's guide for your charger.) |
| | Is the hearing aid charger charged or plugged into a power source? | Charge the hearing aid charger. (See the user's guide for your charger.) |

Regulatory information

Warranties and repairs

The manufacturer provides a warranty on hearing aids in the event of defects in workmanship or materials, as described in applicable warranty documentation. In its service policy, the manufacturer pledges to secure functionality at least equivalent to the original hearing aid. As a signatory to the United Nations Global Compact initiative, the manufacturer is committed to doing this in line with environment-friendly best practices. Hearing aids therefore, at the manufacturer's discretion, may be replaced by new products or products manufactured from new or serviceable used parts, or repaired using new or refurbished replacement parts. The warranty period of hearing aids is designated on your warranty card, which is provided by your hearing care professional.

For hearing aids that require service, please contact your hearing care professional for assistance.

Hearing aids that malfunction must be repaired by a qualified technician. Do not attempt to open the case of hearing aids, as this will invalidate the warranty.

Ambient conditions

Temperature test, operation, transport, and storage information

Our hearing aids are subjected to various tests in temperature and damp heating cycling between -25 °C (-13 °F) and +70 °C (+158 °F) according to internal and industry standards.

During use

During normal operation the temperature should not exceed the limit values of +5 °C (+41 °F) to +40 °C (104 °F) at a relative humidity of 90%, non-condensing. An atmospheric pressure between 700 hPa and 1060 hPa is appropriate.



CAUTION: During use, your hearing aids may reach temperatures up to 43 °C (109 °F).

During transport or storage

During transport or storage, the temperature should not exceed the limit values of:

- -25 °C (-13°F) to +5 °C (41°F)
- +5 °C (41 °F) to +35 °C (95 °F) at a relative humidity up to 90 %, non-condensing
- >+35 °C (95 °F) to +70 °C (158 °F) at a water vapor pressure up to 50 hPa.

Warm-up time: 5 minutes. Cool-down time: 5 minutes

Expected service lifetime

The expected service lifetime for the product when used as intended is:

| Product | Lifetime |
|---|----------|
| Hearing aid | 5 years |
| Built-in rechargeable battery | 5 years |
| Dome | 3 months |
| Receivers | 2 years |
| Electronic accessories (e.g., wireless accessories) | 5 years |

Non-clinical testing (US only)

The devices covered within this user guide have undergone tests for the relevant non-clinical performance testing and biological endpoints in accordance with standards identified below:

- Electrical safety testing is performed according to IEC 60601-1:2005 + A1:2012+A2:2020, IEC 60601-2-66:2019, IEC 60601-1-11 Edition 2.0 2015-1, IEC 62133-2 Edition 1.0 2017- 02 and IEC 62368-1:2018/COR1:2020.
- Electromagnetic compatibility (EMC) testing is performed according to IEC 62479:2010, ANSI IEEE C63.19-2019 and IEC 60601-1-2:2014+A1:2020.
- Radio and Telecommunication testing is performed to be in compliance with applicable parts of the FCC rules in title 47 of the CFR.
- Electroacoustic testing is performed according to ANSI/ASA S3.22-2014 and ANSI/CTA 2051:2017

Usability Engineering was performed in compliance with IEC 62366-1:2015

The devices covered in this user guide passed all tests for the relevant non-clinical performance testing and biological endpoints, namely cytotoxicity (ISO 10993-05:2009), sensitization, and intracutaneous reactivity (ISO 10993-10:2010).

Similarly, usability testing and software verification and validation demonstrated mitigation of risks to an acceptable level as well as reasonable assurance of safe and effective device performance.

Clinical data (US only)

The devices have been evaluated clinically through equivalence, and the devices have been compared to equivalent devices and similar devices on the market with similar intended purpose, e.g., to compensate for hearing impairment by amplifying and transmitting sound to the ear.

Based on technical and clinical data presented for the device in question, the equivalent predecessor and generally similar devices, it is concluded to support the clinical performance expressed in user needs and claims.

The clinical data leaves no questions open regarding clinical performance and is for this reason deemed sufficient

Statement

This device complies with part 15 of the FCC rules and ISED rules. Its operation is subject to the following two conditions:

- 1. This device may not cause harmful interference.
- 2. This device must accept any interference received, including interference that may cause undesired operation.



NOTE: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC rules and ISED rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used according to the instructions, may cause harmful interference to radio communications

However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Redirect or relocate the receiving antenna.
- Increase the separation between the equipment and the receiver
- Connect the equipment to an outlet or a circuit that is different from the one to which the receiver is connected
- Consult the dealer or an experienced radio/TV technician for help.

Changes or modifications can void the user's authority to operate the equipment.

The products are in compliance with the following regulatory requirements:

- In the EU: The device conforms to the Essential Requirements according to Annex I of Council Directive 93/42/EEC for medical devices (MDD).
- GN ReSound A/S hereby declares that the radio equipment types CAR46A, CAR12A, and CAR13A are in compliance with Directive 2014/53/EU.
- The full text of the EU declaration of conformity is available at the following internet address: www.declarations.resound.com
- In the US: FCC CFR 47 Part 15, subpart C.
- Other identified applicable international regulatory requirements in countries outside the EU and US. Please refer to local country requirements for these areas.
- In Canada: these hearing aids are certified under the rules of ISED.
- Japanese Radio Law and Japanese Telecommunications Business Law Compliance: This device has been certified pursuant to the Japanese Radio Law (電波法) and the Japanese Telecommunications Business Law (電気通信事業法). This device should not be modified (otherwise the granted designation number will become invalid).

Type designations

Hearing aid type designations for models included in this user guide are:

CAR46A. FCC ID: X26CAR46A. IC: 6941C-CAR46A CAR12A. FCC ID: X26CAR12A. IC: 6941C-CAR12A CAR13A. FCC ID: X26CAR13A. IC: 6941C-CAR13A.

This device transmits and receives RF signals in the frequency range of 2.4 GHz - 2.48 GHz.

Hearing aid variants

Mini Receiver In-the-Ear (RIE) hearing aids of type CAR46A with FCC ID X26CAR46A. IC number 6941C-CAR46A, and rechargeable Li-lon battery are available in the following variants:

RT961-DRWC, RT761-DRWC, RT561-DRWC Nominal RF output power transmitted is -11.5 dBm.

Mini Receiver In-the-Ear (RIE) hearing aids of type CAR12A with FCC ID X26CAR12A, IC number 6941C-CAR12A, and size 312 Zinc-air battery are available in the following variants:

RT961-DRW. RT761-DRW. RT561-DRW Nominal RF output power transmitted is -10.34 dBm.

Receiver-in-the-ear (RIE) hearing aids of type CAR13A with FCC ID X26CAR13A. IC number 6941C-CAR13A. and size 13 Zinc-air battery are available in the following variants:

RT962-DRW. RT762-DRW. RT562-DRW Nominal RF output power transmitted is -7.1 dBm.

All of the above-mentioned hearing aids contain a magnetic induction radio operating at the 10.66 MHz frequency. The magnetic field strength of the radio is: Max. -24 dBuA/m at a 10 m distance.

Symbols



WARNING: Points out a situation that could lead to serious injuries.



Legal manufacturer.



CAUTION: Indicates a situation that could lead to minor and moderate iniuries.



By prescription only (US).



Follow instructions for use

NOTE: Advice and tips on how to handle



Complies with ACMA requirements.



Equipment includes an RF transmitter.



Complies with IMDA requirements.



Product is a Type B applied part.



Unique Device Identification.



Date of manufacture

vour hearing aid better.



Do not dispose of your hearing aids and batteries with ordinary household garbage. Your hearing aids and batteries should be disposed of at sites intended for electronic waste or returned to your hearing care professional for safe disposal.

Please ask your local hearing care professional about the disposal of your hearing aid.

NOTE: There may be specific regulations in your country.

Technical specifications

RIE - LP/MP receiver

Models: RT962-DRW, RT961-DRW, RT961-DRWC, RT762-DRW, RT761-DRWC, RT762-DRW, RT561-DRWC, RT562-DRW, RT561-DRWC, RT762-DRW, RT561-DRWC, RT762-DRW, RT561-DRWC, RT762-DRW, RT561-DRWC, RT762-DRW, RT561-DRWC, RT762-DRWC, RT762-D DRWC

| | | LP | MP | | |
|---|-----------------------------|-------------------|-------------------|--------|------------------------------------|
| Reference test gain (60 dB SPL input) (The 2 lower curves in the Full-On and Ref.Test.Gain chart) | HFA | 32 | 37 | dB | Maximum Output (OSPL90) |
| Full-on gain (50 dB SPL input) (The 2 upper curves in the Full-On and Ref.Test.Gain chart) | Max HFA | 53 46 | 58 52 | dB | 120 |
| Maximum output (90 dB SPL input) | Max. HFA | 113 108 | 116 114 | dB SPL | 90 |
| Total harmonic distortion | 500 Hz 800 Hz 1600 Hz | 0.5 0.2 0.4 | 0.6 0.6 0.7 | % | 70 1000 10000 10000 Frequency (Hz) |
| Telecoil sensitivity (1 mA/m input)* HFA - SPLIV @ 31.6 mA/m (ANSI) Full-on telecoil sensitivity @ 1 mA/m | Max HFA HFA | 83 92 76 | 90 97 83 | dB SPL | Full-On and Reference Test Gain |
| Equivalent input noise, w/o noise reduction 1/3 Octave Equivalent input noise, w/o noise reduction | 1600 Hz | 20 7 | 20 7 | dB SPL | 60 50 40 |
| Frequency range IEC 60118-0: 2015 | | 100-9410 | 100-9160 | Hz | 30 |
| Battery lifetime (Battery type Rechargeable)** | | 30 | 30 | Hours | 20 |
| Current Drain (Quiescent/Operating) (DRW models only) | | 0.81/1.04 | 0.81/1.04 | mA | 1000 10000 10000 Frequency (Hz) |

Measured according to ANSI S3.22-2014, IEC 60118-0:2015, JIS C 5512; 2015, 2cc coupler.

^{*} Telecoil is only for these models: RT962-DRW, RT762-DRW, RT562-DRW


^{**} The expected operating time of the rechargeable battery depends on active features, the use of wireless accessories, hearing loss, battery age and sound environment.


RIE — LP/MP receiver (US only)

Additional technical data

| Latency, mid frequency delay (2 kHz) | 5.1 | ms |
|---|---------|----|
| Battery current drain, LP (Battery type Rechargeable) | 0.5 | mA |
| Battery current drain, MP (Battery type Rechargeable) | 0.5 | mA |
| Attack/release time (2 kHz syllabic) | 12 / 70 | ms |

Input/Output response, measured in a 2cc coupler at the reference test gain @ 2 KHz





RIE — HP/UP receiver

Models: RT962-DRW, RT961-DRW, RT961-DRWC, RT762-DRW, RT761-DRWC, RT562-DRW, RT561-DRWC, RT561-DRWC

| | | HP | UP | | |
|--|-----------------------------|----------------------------|------------------------|--------|---------------------------------|
| Reference test gain (60 dB SPL input) (The 2 gray curves in the Full-On and Ref.Test.Gain chart) | HFA | 40 | 47 | dB | Maximum Output (OSPL90) |
| Full-on gain (50 dB SPL input) (The 2 upper curves in the Full-On and Ref.Test.Gain chart) | Max. HFA | 65 59 | 75 65 | dB | 120 |
| Maximum output (90 dB SPL input) | Max. HFA | 122 117 | 130 124 | dB SPL | 90 |
| Total harmonic distortion | 500 Hz 800 Hz 1600 Hz | 0.3 0.7 0.7 | 1.0 1.3 0.1 | % | 1000 10000 10000 Frequency (Hz) |
| Telecoil sensitivity (1 mA/m input)* HFA - SPLIV @ 31.6 mA/m (ANS1) Full-on telecoil sensitivity @ 1 mA/m Equivalent input noise, w/o noise reduction 1/3 Octave Equivalent input noise, w/o noise reduction | Max. HFA HFA | 95 100 88 20 7 | 106 108 95 22 | dB SPL | Full-On and Reference Test Gain |
| Frequency range IEC 60118-0: 2015 | | 100-7140 | 100-5010 | Hz | 40 |
| Battery lifetime (Battery type Rechargeable)** | | 30 | 30 | Hours | 20 100 1000 10000 |
| Current Drain (Quiescent/Operating) (DRW models only) | | 0.81/1.04 | 0.81/1.04 | mA | Frequency (Hz) |

Measured according to ANSI S3.22-2014, IEC 60118-0:2015, JIS C 5512: 2015, 2cc coupler.

^{*} Telecoil is only for these models: RT962-DRW, RT762-DRW, RT562-DRW


^{**} The expected operating time of the rechargeable battery depends on active features, the use of wireless accessories, hearing loss, battery age and sound environment.


RIE — HP/UP receiver (US only)

Additional technical data

| Latency, mid frequency delay (2 kHz) | 5.1 | ms |
|---|---------|----|
| Battery current drain, HP (Battery type Rechargeable) | 0.5 | mA |
| Battery current drain, UP (Battery type Rechargeable) | 0.5 | mA |
| Attack/release time (2 kHz syllabic) | 12 / 70 | ms |

Input/Output response, measured in a 2cc coupler at the reference test gain @ 2 KHz





RIE — MM receiver

Models: RT962-DRW, RT961-DRW, RT961-DRWC, RT762-DRW, RT761-DRWC, RT761-DRWC, RT562-DRW, RT561-DRWC.

| | | MM | | |
|---|-----------------------------|-------------------|--------|----------------------------------|
| Reference test gain (60 dB SPL input) (The lower curve in the Full-On and Ref.Test.Gain chart) | HFA | 33 | dB | Maximum Output (OSPL90) |
| Full-on gain (50 dB SPL input) (The upper curve in the Full-On and Ref.Test.Gain chart) | Max. HFA | 53 47 | dB | 120 |
| Maximum output (90 dB SPL input) | Max. HFA | 115 110 | dB SPL | 90 80 |
| Total harmonic distortion | 500 Hz 800 Hz 1600 Hz | 0.1 0.9 0.4 | % | 1000 10000 100000 Frequency (Hz) |
| Telecoil sensitivity (1 mA/m input)* HFA - SPLIV @ 31.6 mA/m (ANSI) Full-on telecoil sensitivity @ 1 mA/m | Max. HFA HFA | 82 93 75 | dB SPL | Full-On and Reference Test Gain |
| Equivalent input noise, w/o noise reduction 1/3 Octave Equivalent input noise, w/o noise reduction | 1600 Hz | 20 7 | dB SPL | 50 |
| Frequency range IEC 60118-0: 2015 | | 100-9100 | Hz | 20 |
| Battery lifetime (Battery type Rechargeable)** | | 28 | Hours | 100 1000 10000 |
| Current Drain (Quiescent/Operating) (DRW models only) | | 0.82/1.04 | mA | Frequency (Hz) |

Measured according to ANSI S3.22-2014, IEC 60118-0:2015, JIS C 5512: 2015, 2cc coupler.

^{*} Telecoil is only for these models: RT962-DRW, RT762-DRW, RT562-DRW


^{**}The expected operating time of the rechargeable battery depends on active features, the use of wireless accessories, hearing loss, battery age and sound environment

RIE — MM receiver (US only)

Additional technical data

| Latency, mid frequency delay (2 kHz) | | ms |
|--|---------|----|
| Battery current drain, MM receiver (Battery type Rechargeable) | | mA |
| Attack/release time (2 kHz syllabic) | 12 / 70 | ms |

Input/Output response, measured in a 2cc coupler at the reference test gain @ 2 KHz



Additional information

Acknowledgements

Portions of this software are written by Kenneth MacKay (micro-ecc) and licensed under the following terms and conditions:

Copyright @ 2014. Kenneth MacKay. All rights reserved.

Redistribution and use in source and binary forms, with or without modification, are permitted provided that the following conditions are met:

- Redistributions of source code must retain the above copyright notice, this list of conditions and the following disclaimer
- Redistributions in binary form must reproduce the above copyright notice, this list of conditions and the following disclaimer in the documentation and/or other materials provided with the distribution.

THIS SOFTWARE IS PROVIDED BY THE COPYRIGHT HOLDERS AND CONTRIBUTORS "AS IS" AND ANY EXPRESS OR IMPLIED WARRANTIES. INCLUDING. BUT NOT LIMITED TO. THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE DISCLAIMED. IN NO EVENT SHALL THE COPYRIGHT HOLDER OR CONTRIBUTORS BE LIABLE FOR ANY DIRECT. INDIRECT. INCIDENTAL, SPECIAL, EXEMPLARY, OR CONSEQUENTIAL DAMAGES (INCLUDING, BUT NOT LIMITED TO. PROCUREMENT OF SUBSTITUTE GOODS OR SERVICES: LOSS OF USE, DATA, OR PROFITS: OR BUSINESS INTERRUPTION). HOWEVER CAUSED AND ON ANY THEORY OF LIABILITY. WHETHER IN CONTRACT, STRICT LIABILITY, OR TORT (INCLUDING NEGLIGENCE OR OTHERWISE) ARISING IN ANY WAY OUT OF THE USE OF THIS SOFTWARE. EVEN IF ADVISED OF THE POSSIBILITY OF SUCH DAMAGE



NOTE: Use of the Made for Apple badge means that an accessory has been designed to connect specifically to iPhone, iPad, and iPod touch models, and has been certified by the developer to meet Apple performance standards. Apple is not responsible for the operation of this device or its compliance with safety and regulatory standards.

© 2023 GN Hearing A/S. All rights reserved. ReSound is a trademark of GN Hearing A/S. Apple. the Apple logo. iPhone, iPad, iPod touch are trademarks of Apple Inc., registered in the U.S. and other countries. App Store is a service mark of Apple Inc., registered in the US and other countries. Android, Google Play and the Google Play logo are trademarks of Google LLC. The Bluetooth word mark and logos are registered trademarks owned by Bluetooth SIG. Inc.

Notes

Notes



Manufacturer according to EU Medical Device Directive 93/42/EEC: Manufacturer according to FDA:

Manufacturer according to Health Canada:

GN ReSound A/S Lautrupbjerg 7 DK-2750 Ballerup Denmark GN ReSound North America 8001 E Bloomington Freeway Bloomington, MN 55420 USA 1-888-735-4327 resound com ReSound Canada 2 East Beaver Creek Road, Building 3 Richmond Hill, ON L4B 2N3 Canada 1-888-737-6863 resound com

ReSound Government Services 8001 E Bloomington Freeway Bloomington, MN 55420 USA 1-800-392-9932 resound.com/veterans (Date of issuance: 2023.03.01) 401402011US_CA_VA- 23.03 - Rev. C