

Cable and Wireless Fitting Guide

ReSound Smart Fit™  20310800 (Noahlink Wireless)	 9022 907 69019 (HI-PRO CABLE/RIGHT/CS44) 9022 907 69029 (HI-PRO CABLE/LEFT/CS44) 3122 150 73781 (FLEX CONNECTOR CLIP CS63) 19181700 (STRIP,FLEX,CS63,PRG)	 15206700 (Programming cable CS63, Right) 15206701 (Programming cable CS63, Left) 19181700 (STRIP,FLEX,CS63,PRG)	 9022 907 69019 (HI-PRO CABLE/RIGHT/CS44) 9022 907 69029 (HI-PRO CABLE/LEFT/CS44) 19533900 (ADAPTER,FLEX,3 PIN,CS53)	 9022 907 69019 (HI-PRO CABLE/RIGHT/CS44) 9022 907 69029 (HI-PRO CABLE/LEFT/CS44) 19902500 (PROGRAMMING ADAPTOR 2)	 18523700 (IIC 10A PROG PILL CABLE, LEFT) 18523800 (IIC 10A PROG PILL CABLE, RIGHT)
ReSound LiNX 3D™					
RIE WL - LTx62	•				•
RIE WL - LTx61	•	•	•	•	
BTE WL - LTx77, LTx88	•				•
BTE WL - LTx67	•	•	•	•	
Custom WL	•	•	•	•	
Custom non-WL		•	•	•	
Custom IIC					•

Please note:

Products supported in ReSound Smart Fit do not support the Airlink dongle.

All ReSound hearing instruments for both wireless and non-wireless **fittings require a fresh battery when fitted with ReSound Smart Fit 1.0.**